

Maintenance Allocation Chart

4110-01-287-7111

Refrigerator, Solid State, Biological, Model, RCB42P

[C – Operator or Crew; O – Unit Maintenance; DS – Direct Support Maintenance; GS – General Support Maintenance; D – Depot Maintenance]

(1) GROUP NUMBER	(2) ASSEMBLY GROUP	(3) MAINTENANCE FUNCTION	(4) MAINTENANCE LEVEL	(5) TOOLS AND EQUIPMENT	(6) REMARKS
00	Refrigerator	Inspect	C		Fluke 80TK Temp Probe or equivalent needed.
		Test	C	19	
		Service	C		
		Safety	O	29	
		Repair	O	01, 02, 04	
		Overhaul	O	01, 02, 04	
		Rebuild	D	01, 02, 03, 04	
01	ACCU	Inspect	O	01	Fluke 80TK Temp Probe or equivalent needed.
		Test	O		
		Service	O	01	
		Repair	O	01	
		Replace	O	01	
02	Thermoelectric Module	Inspect	O	01	
		Test	O	01	
		Service	O	01	
		Repair	O	01, 02	
		Replace	O	01, 02	
03	Fans	Inspect	O	01	
		Test	O	01, 20, 19	
		Service	O	01	
		Repair	O	01, 02	
		Replace	O	01, 02	